



Centro Nacional de Metrología

SIM Regional Supplementary Comparison SIM.L-S1:2007

Calibration of Gauge Blocks by Mechanical Comparison

DRAFT B

July 2012

- **M. Viliesid**
CENAM, Centro Nacional de Metrología
- **C. Colín**
CENAM, Centro Nacional de Metrología
- **T. Chávez**
CENAM, Centro Nacional de Metrología
- **K.P. Chaudhary**
NPLI, National Physical Laboratory INDIA
- **F. Dvořáček**
CMI, Czech Metrology Institute
- **J. Stoup**
NIST, National Institute of Standards and Technology
- **W. Santos**
INMETRO, Instituto Nacional de Metrología
- **L. Vaudagna**
INTI, Instituto Nacional de Tecnología Industrial
- **R. Morales**
DICTUC, Laboratorio Nacional de Longitud
- **A. Acquarone**
LATU, Laboratorio Tecnológico del Uruguay
- **J. Carrasco**
INDECOPI, Servicio Nacional de Metrología
- **M. Vega**
IBMETRO, Instituto Boliviano de Metrología
- **M. Salazar**
INEN, Instituto Ecuatoriano de Normalización
- **V. Gil**
SIC, Superintendencia de Industria y Comercio
- **J. Dimas**
CENAMEP, Centro Nacional de Metrología de Panamá AIP
- **E. Reyes**
LACOMET, Laboratorio Costarricense de Metrología
- **S. Durga**
BSJ, Bureau of Standards Jamaica
- **T. Burton**
BSJ, Bureau of Standards Jamaica

Centro Nacional de Metrología (CENAM)
km 4.5 Carretera a Los Cués,
El Marqués, Querétaro
76246, MEXICO

1. Introduction

The Mutual Recognition Arrangement (MRA) of the *Conférence Internationale des Poids et Mesures (CIPM)* signed by the National Metrology Institutes (NMI) of different nations provides mutual recognition among the NMI of their national standards and their calibration services. A database has been set up by the *Bureau International des Poids et Mesures (BIPM)* at its website where the Calibration and Measurement Capabilities (CMC) of each NMI are posted. To support the CMC claims of the NMI, the MRA requires, among other things, that they participate on a regular basis in Key Comparisons (KC) that test key measuring techniques. This would prove their technical competence, that they can provide this calibration service with the claimed uncertainty of the corresponding CMC and that they have metrological equivalence with the other signatory NMI that provide the same service.

The CIPM has therefore instructed the different CC to identify key techniques in order to define KC, as it is, for example, the calibration of Gauge Blocks (GB) by optical interferometry, identified as a key measuring technique by CCL. Additionally, the CC as well as the regions may also identify other important comparisons called supplementary and identified with a S. The SIM region has identified the Calibration of GB by Mechanical Comparison as one of these comparisons. The *Centro Nacional de Metrología (CENAM)* was designated by SIM as pilot laboratory and CENAM has carried out this exercise under the name of SIM.L-S1:2007.

The calibration of GB by Mechanical Comparison is indeed a technique of paramount importance as it is at the highest level in the traceability chain of length for most countries of the Americas. This comparison is meant to support the submitted CMC of these countries.

The mesurand is the central length of the GB as defined in [1] and the circulated GB were used for two comparisons carried out in two stages:

- First stage, SIM.L- K1:2007, Calibration of GB by Optical Interferometry. Circulation from 2007-11-01 to 2010-04-25. The GB were also measured by Mechanical Comparison for those NMI also participating in SIM.L-S1:2007.
- Second stage, SIM.L-S1:2007, Calibration of GB by Mechanical Comparison. Circulation from 2010-03-02 to 2011-05-06 for the participants that measured only by Mechanical Comparison, this dates included two control measurements by interferometry by the pilot laboratory.

In this second comparison there were 16 participants, 14 from the Americas, and 2 invited NMI from other regions. The circulation in the second stage had 10 participants. It should be noted that the circulation took relatively short time for the large number of participants thanks to the hand delivery from one NMI to the following one in nine out of the 10 steps so that customs clearance was simplified or avoided.

2. Participants

This comparison had 16 participants. **Table 1** shows the participating NMI and their corresponding contact person and information.

Contact	NMI	Information
Carlos Colín Castellanos Trinidad Chávez	CENAM , Centro Nacional de Metrología km 4.5 Carretera a los Cués, El Marqués CP 76241, Querétaro, MÉXICO .	Tel. +52 442 211 0500 Fax +52 442 211 0577 e-mail: ccolin@cenam.mx
K. P. Chaudhary	NPLI , National Physical Laboratory INDIA Dr. K.S. Krishnan Road, New Delhi 110012, INDIA .	Tel. +91 11 25732865 Fax +91 11 25726938 e-mail: kpc@mail.nplindia.ernet.in
Ing. Vladimír Stezka Ing. Frantisek Dvořáček	CMI , Czech Metrology Institute Slunecna 23 460 01 Liberec CZECH REPUBLIC .	Tel. +42 485 107 532 Fax +42 485 104 466 e-mail: vstezka@cmi.cz fdvoracek@cmi.cz
Leandro Vaudagna	INTI , Instituto Nacional de Tecnología Industrial División de Metrología Dimensional-Rafaela km 227,6 Ruta Nac. N° 34 CP 2300, Rafaela, Santa Fé, ARGENTINA	Tel. +54 34 92440471 Fax +54 34 92422804 e-mail: vaudagna@inti.gov.ar
Wellington Santos Barros	INMETRO , Instituto Nacional de Metrologia, Normalização e Qualidade Industrial. Av. N.Sra. das Graças, 50 – Villa Operária – Xerém – Duque de Caixas – RJ. CEP 25250- 020, BRASIL .	Tel. +55 21 2679-9271 Fax +55 21 2679-9207 e-mail: wsbarros@inmetro.go.br
John Stoup	NIST , National Institute of Standards and Technology Room B113, Metrology Building Gaithersburg, MD 20899-0001 USA	Tel. +1 301 975 3476 Fax + 1 301 869 0822 e-mail: John.Stoup@nist.gov
Roberto Morales	DICTUC , Laboratorio Nacional de Longitud Avenida Vicuña Mackenna 4860 – Macul – Santiago – (edificio n° 9 metrología), CHILE .	Tel. 56 2 3544624 Fax 56 2 3544624 e-mail: metrologia@dictuc.cl
Alejandro Acquarone Luis Mussio	LATU , Laboratorio Tecnológico del Uruguay Avenida Italia 6201, Montevideo, URUGUAY . CP 11500	Tel. 598 2 601 3724 ext 298 Fax 598 2 601 8554 e-mail: lmussio@latu.org.uy aacqua@latu.org.uy
Janet Carrasco Tuesta	INDECOPI , Servicio Nacional de Metrología Calle de la Prosa 138, San Borja Lima 4,1 PERÚ .	Tel. 51 1 224 7800 1618 Fax 51 1 224 7800 1264 e-mail: jcarrasco@indecopi.gob.pe
María del Carmen Vega Amonzabel	IBMETRO , Instituto Boliviano de Metrología Av. Camacho No. 1488 La Paz, BOLIVIA .	Tel. 591 2 2372046 Fax 591 2 2310037 e-mail: mvega@ibmetro.gob.bo
Manuel Salazar	INEN , Instituto Ecuatoriano de Normalización Casilla: 17-01-3999 Quito – ECUADOR .	Tel. 593 2 2343716 Fax 593 2 2344394 e-mail: msalazar@inen.gov.ec
Victor Hugo Gil	SIC , Superintendencia de Industria y Comercio División de Metrología COLOMBIA .	Tel. 571 3153265 - 69 Fax 571 3153292 e-mail: vgil@correo.sic.co
Julio Dimas	CENAMEP , Centro Nacional de Metrología de Panamá AIP, Edificio 215, Ciudad del Saber, PANAMÁ .	Tel. 507 – 517-0081 Fax 507 – 517-0019 e-mail: jdimas@cenamep.org.pa
Eduardo Reyes	LACOMET , Laboratorio Costarricense de Metrología. Apartado Postal 1736 – 2050, San Pedro de Montes de Oca, San Jose. COSTA RICA .	Tel. 506 - 2283 - 6580 Fax 506 - 2283 - 5133 e-mail: ereyes@lacomet.go.cr
Theodore Reddock Francis Hamilton	TTBS , Trinidad and Tobago Bureau of Standards Century Drive Trincity Industrial Estate Macoya, Tunapuna, TRINIDAD AND TOBAGO .	Tel. 868-662-8827 Fax 868-663-4335 e-mail: Theodore.Reddock@ttbs.org.tt e-mail: Francis.Hamilton@ttbs.org.tt
Tomokie Burton Carlton Thomas Siew Durga	BSJ , Bureau of Standards, Jamaica 6 Winchester Rd., Kingston 10. JAMAICA .	Tel.: 926-3140-5 ext. 1102 Fax: 929-4736 e-mail: cthomas@bsj.org.jm e-mail: TBurton@bsj.org.jm

Tabla 1. List of participants in comparison SIM.L-S1:2007.

3. Circulation Schedule

As we mentioned, nine out of the ten participants in the second circulation stage delivered the artifacts by hand to the following participant which made the circulation time relatively short. Table 2 shows the circulation schedule for the two circulation stages. *LACOMET* retained 13 weeks the GB because of end-of-the-year holidays and internal administrative problems, they mentioned.

NMI	Dates		Dates
	Reception	Reception	
Stage One Circulation			
NPLI	2008-03-02	2008-06-09	2009-08-07
CMI	2008-06-16	2008-08-05	2008-10-17
NIST	2009-04-20	2009-07-15	2010-03-02
INMETRO	2009-08-11	2009-09-08	2009-09-21
INTI	2009-11-13	2010-01-11	2010-01-11
Stage Two Circulation			
DICTUC	2010-04-29	2010-05-18	2010-06-15
LATU	2010-06-11	2010-07-06	2010-07-21
INDECOPI	2010-07-06	2010-07-30	2010-08-12
IBMETRO	2010-08-02	2010-08-18	2010-09-24
INEN	2010-08-18	2010-09-13	2011-05-16
SIC	2010-09-13	2010-10-13	2010-11-10
CENAMEP	2010-10-13	2010-11-09	2011-01-28
LACOMET	2010-11-09	2011-02-09	2011-04-01
TTBS	2011-02-14	2011-03-16	2011-04-13
BSJ	2011-03-16	2011-04-27	2011-06-09
CENAM (Pilot)	2011-04-28		2011-05-25

Table 2. SIM.L-S1:2007 dates of reception and shipment of artifacts and reception of results by the pilot laboratory.

4. Comparison Artifacts

A total of 14 grade K (according to [1]) rectangular GB were selected for the exercise. Seven steel GB and seven ceramics GB covering the range of short GB (from 0.5 mm to 100 mm). The specifications on the GB are shown in **tables 3** and **4**. The associated Coefficients of Thermal Expansion (CET) shown in the tables are those quoted by the manufacturer.

Nominal Length (mm)	Serial Number	Coefficient of Thermal Expansion (10^{-6} K^{-1})	Manufacturer
1.000 5	010223	10.9 ± 1	Mitutoyo
5	000482	10.9 ± 1	Mitutoyo
7	010764	10.9 ± 1	Mitutoyo
10	001329	10.9 ± 1	Mitutoyo
50	012254	10.9 ± 1	Mitutoyo
75	010630	10.9 ± 1	Mitutoyo
100	010850	10.9 ± 1	Mitutoyo

Table 3. Steel Gauge Blocks.

Nominal Length (mm)	Serial Number	Coefficient of Thermal Expansion (10^{-6} K^{-1})	Manufacturer
1.000 5	000288	9.3 ± 1	Mitutoyo
5	051836	9.3 ± 1	Mitutoyo
7	010323	9.3 ± 1	Mitutoyo
10	052351	9.3 ± 1	Mitutoyo
50	011002	9.3 ± 1	Mitutoyo
75	010370	9.3 ± 1	Mitutoyo
100	010773	9.3 ± 1	Mitutoyo

Table 4. Ceramics Gauge Blocks.

5. Measurement Protocol

Detailed instructions were included in the technical protocol. Participants were invited to perform the measurements according to their own calibration procedures, used to calibrate the GB of their customers.

The measurement was performed in all cases with a double probe GB comparator and in the vertical position as indicated in [1]. The method determines the difference in central length, l_c , of two GB of same nominal length set beside in the comparator platen as illustrated in **Figure 1**. The first GB is the laboratory's reference GB, calibrated by optical interferometry; and the test GB which is under circulation.




Figure 1. Illustration of the calibration method of GB by mechanical comparison showing the different variables of influence (taken from [9]).

6. Measuring Instruments

All participants measured with double probe electromechanical GB comparators with a resolution of 10 nm. Table 5 shows the makes, models and characteristics of the different instruments.

NMI	Manufacturer	Model	Measuring range mm	Traceability	Temperature variation range during measurements (°C)
CENAM	TESA	TESA-UPC	0 - 102	To SI standards of CENAM via GB calibrated by interferometry	19.32 – 19.60
NPLI	Mahr	Not specified	0 – 175	Not specified	19.5 – 20.5
CMI	TESA	TESA-UPC	0 – 100	To the Czech National Standard of Length (He-Ne/I2 633nm, He-Ne/I2 543.5nm, fs comb)	Not specified
NIST	Mahr/Federal	130B-24	0 -102	NIST maintained Iodine-Stabilized Laser	20.08 – 20.17
INMETRO	TESA	TESA-UPC	0 - 102	To SI standards of INMETRO via GB calibrated by interferometry	20.0 – 20.5
INTI	Mahr	826E	0 - 175	To SI standards of INTI via GB calibrated by interferometry	20 ± 0.5
DICTUC	TESA	TESA-UPC	0 - 100	To SI standards of PTB via GB calibrated by interferometry	Steel 19.91 – 20.05 Ceramics 19.85 – 20.16
LATU	Mahr	826	0 - 100	Comparator to SI standards of PTB via GB calibrated by interferometry and to SI standards of CENAM via GB calibrated by interferometry	19.68 – 20.50
INDECOPI	Mahr	826 PC	0 – 175	To SI standards of CENAM via GB calibrated by interferometry	20.0 ± 0.5
IBMETRO	Steinmayer	EMP II	0 – 100	To SI standards of PTB via GB grade 0	19.95 – 20.15
INEN	Mahr	826	0 - 170	Not indicated, only mentioned that their GB are calibrated by interferometry	19 – 21
SIC	TESA	UPD	0 - 500	To SI standards of METAS via GB calibrated by interferometry	20 – 20.3
CENAMEP	Mitutoyo	GBCD-250	0 - 250	To SI standards of CENAM via GB calibrated by interferometry	20.3 – 20.6
LACOMET	TESA	TESA-UPC	0.5 - 100	To SI standards of PTB and CENAM via GB calibrated by interferometry	0.34
TTBS	TESA	TESA-UPC	0.5 - 100	To SI standards of METAS and NPL via GB calibrated by interferometry	0.018
JBS	Mahr-Federal	2247386	1 - 100	To SI standards of NIST	19.60 – 20.20

Tabla 5. GB comparators, measurement range, traceability and temperature variation of the participant laboratories.

7. State and Behavior of Artifacts


7.1 State of the Artifacts upon Reception

The participants were to inspect the state of the artifacts upon reception and inform the pilot according to the protocol. Although the selected GB were not brand new, they were in good conditions. They suffered some damage after the circulation, but the stability and

the results obtained in the comparison prove the damages did not hamper or alter the measurements and the pilot laboratory was able to bring them all to a measurement platen without problems after circulation. **Figures 1 through 11** show the physical conditions of some of the steel GB upon reception at the pilot laboratory at the end of the exercise. The steel GB ended-up with quite a few scratches and specifically, the 5 mm GB, also presented rust spots.



Figures 1, 2 and 3. Aspect of the measuring faces of the 1.0005, 5 and 7 mm Steel GB at the end of circulation.



Figures 4, 5 and 6. Aspect of the measuring faces of the 10, 50 and 75 mm Steel GB at the end of circulation.



Figure 7. Aspect of the measuring faces of the 100 mm Steel GB at the end of circulation.

Four out of the seven ceramic GB suffered some damage on one of their measuring faces as shown in **Figures 8 to 11**, which consisted of burs or chipped edges. However, this condition did not hamper the wringing or caused any variation in length as it was proved by the control measurements performed at the end by the pilot laboratory by interferometry as well as by mechanical comparison.



Figures 8, 9, 10 and 11. Aspect of the measuring faces of the 10, 50, 75 and 100 mm Ceramic GB at the end of circulation.

7.2 Stability of the Standards

The GB were measured by interferometry several times by the pilot laboratory to verify their stability: when they were purchased (2002), two years before starting the comparison (Nov. 2005), before circulating them (Nov. 2007), after the round by interferometry (April 2010) and at the end of the circulation (May 2011). **Table 6** shows the deviations from nominal length determined at these different occasions for the steel GB, including the stated values on the certificates of the manufacturer. **Graphs 1** through **7** show these values for each GB.

Serial Number	Nominal length (mm)	Deviation from nominal value (nm)					
		Manufacturer certificate 2001	2002	2005	2007	2010	2011
010223	1.000 5	0	5	3	-4	-9	-2
000482	5	40	14	11	35	20	27
010764	7	30	19	13	-5	1	-10
001329	10	50	31	22	37	21	31
012254	50	60	46	3	7	-3	8
010630	75	-50	-54	-104	-100	-107	-106
010850	100	20	18	-50	-51	-64	-50

Table 6. Pilot Laboratory measured values of the steel GB at different occasions.






Table 7 shows the deviations from nominal length determined at these different occasions for the ceramics GB, including the stated values on the certificates of the manufacturer. **Graphs 8** through **14** show these values for each GB.

Serial Number	Nominal length (mm)	Deviation from nominal value (nm)					
		Manufacturer certificate 2001	2002	Manufacturer certificate 2005	2007	2010	2011
000288	1.0005	0	-6	----	-14	7	-17
051836	5	----	----	13	12	8	14
010323	7	50	46	----	57	48	35
052351	10	----	----	3	-13	-19	-29
011002	50	90	95	----	139	117	97
010370	75	100	110	----	118	124	123
010773	100	-60	-42	----	-34	-36	-41

Table 7. Pilot Laboratory measured values of the ceramics GB in different occasions.





8. Measurement Results of Participants

All laboratories sent their results by e-mail. All information was received on the specified formats from appendices A, B, C, D and E of the Technical Protocol.

8.1 Measurement of the Central Length

Tables 8 and 9 and graphs 16 through 22, show the deviations of the central length with respect to nominal values and the claimed standard measurement uncertainties of each participant for the seven steel GB. Additionally, graph 15 shows the claimed standard uncertainties of all participants.

Nominal Value mm	Deviation (e_{ij}) from nominal length for Steel GB nm							
	NPLI	CMI	NIST	INMETRO	INTI	LATU	DICTUC	INDECOPI
1.000 5	-20	20	-27	20	-12	-35	-15	20
5	10	60	27	62	-4	4	5	20
7	-40	20	2	21	-38	-12	-35	10
10	10	70	39	24	12	39	20	40
50	-30	80	15	11	-11	13	-10	0
75	60	-50	-107	-93	-127	-126	-110	-80
100	10	40	-67	-60	-66	-68	-60	10

Table 8A. Measurement results of the participants for the Steel GB.

Nominal Value mm	Deviation (e_{ij}) from nominal length for Steel GB nm							
	IBMETRO	INEN	SIC	CENAMEP	LACOMET	TTBS	BSJ	CENAM
1.000 5	-40	-10	-27	-13	22	30	Not reported	-2
5	0	10	-6	-3	46	70	11	23
7	-40	-30	-38	-13	27	20	-37	6
10	20	20	-7	31	66	20	15	24
50	-10	10	0	-10	36	200	-26	14
75	-130	-140	-127	-143	-83	-40	-295	-109
100	-70	40	-23	-112	-68	290	-47	-33


Table 8B. Measurement results of the participants for the Steel GB.



Nominal Value mm	Claimed standard uncertainties, $u(e_{ij})$, for Ceramic GB nm							
	NPLI	CMI	NIST	INMETRO	INTI	LATU	DICTUC	INDECOPI
1.000 5	26	23	13	29.5	23.0	26.0	27	27
5	28	23.1	13	29.8	23.0	26.0	27	27
7	29	23.1	13	29.8	23.0	26.0	27	27
10	30	23.3	13	29.8	23.0	26.0	27	27
50	53	29.2	15	35.2	35.0	38.0	29	36
75	67	35.5	20	41.4	46.0	40.0	34	45
100	81	42.7	30	49.1	58.0	47.0	36	55



Table 9A. Claimed standard uncertainties of the participants for the Steel GB.



Nominal Value mm	Claimed standard uncertainties, $u(e_{ij})$, for Ceramic GB nm							
	IBMETRO	INEN	SIC	CENAMEP	LACOMET	TTBS	BSJ	CENAM
1.000 5	34	55.4	23	12	25.3	52	Not reported	14.5
5	35	57.2	23	12	25.4	53	28	14.7
7	35	58	23	12	36.7	53	29	14.9
10	36	59.3	24	13	36.9	77	31	15.2
50	44	76.5	34	35	43.7	160	71	27.6
75	50	87.3	44	51	43.8	170	100	38.1
100	55	98	55	66	54.4	170	130	49.2


Table 9B. Claimed standard uncertainties of the participants for the Steel GB.











Tables 10 and 11 and graphs 24 through 30, show the deviations of the central length with respect to nominal values and their claimed standard measurement uncertainties of each participant for the seven ceramic GB. Graph 23 shows the claimed standard uncertainties of the participants.

Nominal Value mm	Deviation (e_{ij}) from nominal length for Ceramics GB nm								
	NPLI	CMI	NIST	INMETRO	INTI	LATU	DICTUC	INDECOPI	
1.000 5	-40	0	-27	-32	-9	-37	-17	-15	
5	0	30	14	-14	10	-29	-17	-15	
7	10	50	58	34	35	13	13	25	
10	-80	0	-9	-24	-15	-18	-32	-25	
50	20	140	116	97	89	103	65	55	
75	40	150	148	89	122	111	101	135	
100	-80	10	-5	-55	16	-53	-65	-15	

Table 10A. Measurement results of the participants for Ceramics GB.

Nominal Value mm	Deviation (e_{ij}) from nominal length for Ceramics GB nm								
	IBMETRO	INEN	SIC	CENAMEP	LACOMET	TTBS	BSJ	CENAM	
1.000 5	310	-10	-5	-12	-9	60	Not reported	-10	
5	340	-20	-10	-1	-16	80	-36	-6	
7	370	20	32	38	44	90	-18	65	
10	320	-10	-40	-18	-15	Not reported	-68	-20	
50	460	140	138	103	84	480	-16	100	
75	480	170	155	164	102	480	-140	127	
100	360	40	40	2	8	660	-195	-12	


Table 10B. Measurement results of the participants for Ceramics GB.



Nominal Value mm	Claimed standard uncertainties, $u(e_{ij})$, for Ceramics GB							
	nm							
	NPLI	CMI	NIST	INMETRO	INTI	LATU	DICTUC	INDECOPI
1.000 5	26	23	13	32.5	23	26	27	27
5	28	23.1	13	33.2	23	26	27	27
7	29	23.1	13	32.7	23	26	27	27
10	30	23.3	13	32.7	23	26	27	27
50	53	29.2	15	38.1	33	36	28	38
75	67	35.5	20	44	42	45	35	48
100	81	42.7	30	50.6	52	47	34	59



Table 11A. Claimed standard uncertainties of the participants for Ceramics GB.



Nominal Value mm	Claimed standard uncertainties, $u(e_{ij})$, for Ceramics GB							
	nm							
	IBMETRO	INEN	SIC	CENAMEP	LACOMET	TTBS	BSJ	CENAM
1.000 5	34	55.4	23	12	25.3	58	Not reported	19.0
5	35	57.2	23	12	25.4	58	39	19.2
7	35	58	23	12	25.5	58	40	19.4
10	36	59.3	24	13	25.7	Not reported	43	19.9
50	44	76.5	34	34	43.5	160	99	34.5
75	50	87.3	44	50	43	170	140	47.1
100	56	98	55	66	55.3	180	182	60.6


Table 11B. Claimed standard uncertainties of the participants for Ceramics GB.











9. Reference Value (RV) and equations to determine the performance of participants

All usual parameters of the central tendency were calculated: the median, the simple mean and the inverse-variance weighted mean. All of these values appear in **Annex A**. However, the Supplementary Comparison Reference Values were obtained from the Regional Key Comparison SIM.L-K1:2007, as the artifacts were the same. These RV were calculated as the simple mean for all consistent results of the interferometric comparison [2].

The Reference Values, and their Expanded Uncertainties, for the different GB *j* of both materials are shown in **Table 13**.

Supplementary Comparison Reference Values (RV)				
Nominal Length mm	Steel		Ceramics	
	Ref. Value,		Ref. Value,	
1.005	-9.3	8.2	-3.9	7.8
5	26.4	8.3	10.6	7.9
7	-2.6	8.3	51.7	7.9
10	36.2	7.8	-14.3	8.1
50	4.9	12.5	105.9	11.3
75	-105.6	17.0	136.2	13.9
100	-42.0	17.1	-23.1	15.6

Table 13. Reference values (simple mean of largest sub-set of consistent results of SIM.L-K1:2007 comparison) as deviations from Nominal Value and corresponding Expanded Uncertainty for both steel and ceramic GB.

For each laboratory, i , which measures each gauge block, j , let the measured deviation from nominal length be denoted by d_{ij} and calculated as,

$$d_{ij} = e_{ij} - \bar{e}_j \quad (1)$$

Statistical consistency of the results with their associated uncertainties can be verified by calculating the normalized error E_n .

$$E_n = \frac{|d_{ij}|}{\sqrt{U^2(e_{ij}) + U^2(\bar{e}_j)}} \quad (2)$$

If E_n is greater than 1 it is considered that the result is inconsistent.

10. Participants Results of the Comparison

The reported measurement results were analyzed by simple statistical means to allow identification of any significant bias. **Tables 14** and **15** show the differences of the results of the participants with respect to the **RV** of each GB j , d_{ij} , and the corresponding Normalized Error, E_n calculated from equation (2). Note that the uncertainties in this equation are expanded uncertainties.

NMI (i→)	NPLI		CMI		NIST		INMETRO		INTI		LATU		
	Nominal Length (j)	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n
1.000 5		-10.7	0.20	29.3	0.63	-17.7	0.65	29.3	0.49	-2.7	0.06	-25.7	0.49
5		-16.4	0.29	33.6	0.72	0.6	0.02	35.6	0.59	-30.4	0.65	-22.4	0.43
7		-37.4	0.64	22.6	0.48	4.6	0.17	23.6	0.39	-35.4	0.76	-9.4	0.18
10		-26.2	0.43	33.8	0.72	2.8	0.10	-12.2	0.20	-24.2	0.52	2.8	0.05
50		-34.9	0.33	75.1	1.26	10.1	0.31	6.1	0.09	-15.9	0.22	8.1	0.11
75		165.6	1.23	55.6	0.76	-1.4	0.03	12.6	0.15	-21.4	0.23	-20.4	0.25
100		52	0.32	82	0.94	-25	0.40	-18	0.18	-24	0.20	-26	0.27

Table 14A. Deviation from reference value for each GB, d_{ij} and Normalized Error, E_n , of the **Steel GB** for the first six participants.

NMI (i→)	DICTUC		INDECOPI		IBMETRO		INEN		SIC		CENAMEP		
	Nominal Length (j)	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n
1.000 5		-5.7	0.10	29.3	0.54	-30.7	0.45	-0.7	0.01	-17.7	0.38	-3.7	0.15
5		-21.4	0.39	-6.4	0.12	-26.4	0.37	-16.4	0.14	-32.4	0.69	-29.4	1.16
7		-32.4	0.59	12.6	0.23	-37.4	0.53	-27.4	0.24	-35.4	0.76	-10.4	0.41
10		-16.2	0.30	3.8	0.07	-16.2	0.22	-16.2	0.14	-43.2	0.89	-5.2	0.19
50		-14.9	0.25	-4.9	0.07	-14.9	0.17	5.1	0.03	-4.9	0.07	-14.9	0.21
75		-4.4	0.06	25.6	0.28	-24.4	0.24	-34.4	0.20	-21.4	0.24	-37.4	0.36
100		-18	0.24	52	0.47	-28	0.25	82	0.42	19	0.17	-70	0.53

Table 14B. Deviation from reference value for each GB, d_{ij} and Normalized Error, E_n , of the **Steel GB** for the next six participants.

NMI (i→)	LACOMET		TTBS		BSJ		CENAM	
	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n
Nominal Length (j_i)								
1.000 5	31.3	0.61	39.3	0.38	Not reported	Not reported	7.3	0.24
5	19.6	0.38	43.6	0.41	-15.4	0.27	-3.4	0.11
7	29.6	0.40	22.6	0.21	-34.4	0.59	8.6	0.28
10	29.8	0.40	-16.2	0.11	-21.2	0.34	-12.2	0.39
50	31.1	0.35	195.1	0.61	-30.9	0.22	9.1	0.16
75	22.6	0.25	65.6	0.19	-189.4	0.94	-3.4	0.04
100	-26	0.24	332	0.98	-5	0.02	9	0.09

Table 14C. Deviation from reference value for each GB, d_{ij} and Normalized Error, E_n , of the Steel GB for the last four participants.

NMI (i→)	NPLI		CMI		NIST		INMETRO		INTI		LATU	
	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n
Nominal Length (j_i)												
1.000 5	-36.1	0.69	3.9	0.08	-23.1	0.85	-28.1	0.43	-5.1	0.11	-33.1	0.63
5	-10.6	0.19	19.4	0.41	3.4	0.13	-24.6	0.37	-0.6	0.01	-39.6	0.75
7	-41.7	0.71	-1.7	0.04	6.3	0.23	-17.7	0.27	-16.7	0.36	-38.7	0.74
10	-65.7	1.09	14.3	0.30	5.3	0.19	-9.7	0.15	-0.7	0.01	-3.7	0.07
50	-85.9	0.81	34.1	0.57	10.1	0.32	-8.9	0.12	-16.9	0.25	-2.9	0.04
75	-96.2	0.71	13.8	0.19	11.8	0.28	-47.2	0.53	-14.2	0.17	-25.2	0.28
100	-56.9	0.35	33.1	0.38	18.1	0.29	-31.9	0.31	39.1	0.37	-29.9	0.31

Table 15A. Deviation from reference value for each GB, d_{ij} and Normalized Error, E_n , of the Ceramics GB for the first six participants.

NMI (i→)	DICTUC		INDECOPI		IBMETRO		INEN		SIC		CENAMEP	
	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n
Nominal Length (j_i)												
1.000 5	-13.1	0.24	-11.1	0.20	313.9	4.59	-6.1	0.05	-1.1	0.02	-8.1	0.32
5	-27.6	0.51	-25.6	0.47	329.4	4.68	-30.6	0.27	-20.6	0.44	-11.6	0.46
7	-38.7	0.71	-26.7	0.49	318.3	4.52	-31.7	0.27	-19.7	0.42	-13.7	0.54
10	-17.7	0.32	-10.7	0.20	334.3	4.61	4.3	0.04	-25.7	0.53	-3.7	0.14
50	-40.9	0.72	-50.9	0.66	354.1	3.99	34.1	0.22	32.1	0.47	-2.9	0.04
75	-35.2	0.49	-1.2	0.01	343.8	3.41	33.8	0.19	18.8	0.21	27.8	0.28
100	-41.9	0.60	8.1	0.07	383.1	3.39	63.1	0.32	63.1	0.57	25.1	0.19

Table 15B. Deviation from reference value for each GB, d_{ij} and Normalized Error, E_n , of the Ceramics GB for the next six participants.

NMI (i→)	LACOMET		TTBS		BSJ		CENAM	
	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n	d_{ij}	E_n
Nominal Length (j_i)								
1.000 5	-5.1	0.10	63.9	0.55	Not reported	Not reported	-6.1	0.16
5	-26.6	0.52	69.4	0.60	-46.6	0.59	-16.6	0.42
7	-7.7	0.15	38.3	0.33	-69.7	0.87	13.3	0.34
10	-0.7	0.01	Not reported	Not reported	-53.7	0.62	-5.7	0.14
50	-21.9	0.25	374.1	1.17	-121.9	0.61	-5.9	0.08
75	-34.2	0.39	343.8	1.01	-276.2	0.99	-3.4	0.10
100	31.1	0.28	683.1	1.90	-171.9	0.47	11.1	0.09

Table 15C. Deviation from reference value for each GB, d_{ij} and Normalized Error, E_n , of the Ceramics GB for the last four participants.

Table 16 shows the Root Mean Square (RMS) values of the deviations of the participants, δ_{RMS} , with respect to the RV. It gives a general idea of the deviations of each participant with respect to RV. It is determined as:

$$\delta_{RMS} = \sqrt{\frac{\sum_{i=1}^n (e_{ij} - \bar{e}_j)^2}{n}} \quad (3)$$

Where:

e_{ij} – Deviation from nominal of laboratory i on GB j ,

\bar{e}_j – RV of GB j ,

n – Number of GB

NMI	δ_{RMS}	
	Steel GB	Ceramics GB
NPLI	69.5	62.5
CMI	52.2	20.9
NIST	12.4	13.0
INMETRO	21.9	27.1
INTI	24.1	18.2
LATU	18.6	28.6
DICTUC	18.4	32.5
INDECOPI	25.2	24.7
IBMETRO	26.5	340.3
INEN	36.3	34.4
SIC	27.6	31.3
CENAMEP	32.8	16.1
LACOMET	27.5	22.1
TTBS	149.7	350.0
BSJ	80.8	147.5
CENAM	8.1	10.5

Table 16. RMS Values of the deviations with respect to the RV, δ_{RMS} , of the participants.

In this case, as the RV is determined external to the comparison, it is not possible calculate Birge ratio.

11. Discussion and Conclusions

11.1 Discussion

- The comparison was linked to a previous interferometric comparison that measured the same artifacts. This was an advantage as the RV were obtained from the interferometric stage providing low uncertainty RV for the Mechanical Comparison exercise. Not only were the RV obtained by a metrological superior technique, but it was the result of the measurement of many participants that measured by this technique.
- In the second stage of circulation which included only those NMI that measured exclusively by mechanical comparison (10 laboratories), the timing of circulation (from 2010-04-15 to 2011-04-28) was short thanks to the hand-delivery of the artifacts to the following participant. We would like to thank the participants for having taken this trouble and we suggest adopting this transport option whenever

possible as it reduces the time of circulation and the risk of damage to the artifacts during transport.

- There was also some time saving during circulation in the first stage, as the NMI that participated in both exercises received the artifacts only once and measured by both techniques during the same period.
- Declared standard uncertainties among participants spread over a 6 fold range, going from 10 nm to 60 nm for the shortest GB and from 25 nm to 180 nm for the 100 mm GB.
- The calibration uncertainty is an important issue because it should be small enough to allow discriminating if a GB is within its manufacture tolerance or not. According to [1] there are tolerances on the deviation from nominal length depending on the grade and length of the GB. One of the purposes of the calibration is to verify that a GB is within its tolerance of fabrication. For the purpose of this discussion we dare suggest that the calibration should be performed with an uncertainty not larger than one third of the Limit Deviation of Length from Nominal Value. Based on this argument, **Table 17** suggests the maximum Expanded Permissible Uncertainties for the different GB:

Nominal Length		Grade 0		Grade 1		Grade 2	
mm							
More than	Up to	Limit Deviation of Length from Nominal Length	Maximum Permissible Expanded Uncertainty	Limit Deviation of Length from Nominal Length	Maximum Permissible Expanded Uncertainty	Limit Deviation of Length from Nominal Length	Maximum Permissible Expanded Uncertainty
		nm	nm	nm	nm	nm	nm
0	10	120	40	200	70	450	150
10	25	140	50	300	100	600	200
25	50	200	70	400	130	800	270
50	75	250	80	500	170	1 000	330
75	100	300	100	600	200	1 200	400

Table 17. Suggested Maximum Expanded Uncertainties for the Calibration of different grade GB.

- According to the declared uncertainties, the participant laboratories should review up to what grade of GB they able to calibrate according to the previous discussion.
- A few participants had the same traceability source because their master GB were calibrated at a same laboratory. However, we consider the influence of these correlations minimal and they were not taken in account in the present analysis.

11.2 Conclusions

- From Section 7 we observe that there were no appreciable changes on the measurements performed by the pilot laboratory of the ensemble of the GB of both materials over the last five years. Even though some drift may be appreciated on the steel GB during their first years of their history, the values shown prove they reached stability since 2005 approximately. Therefore, it can be assumed that the artifacts behaved adequately during the comparison exercise and that the exercise was valid.

- It may not be asserted that there was more consistency in one material or another and results were similar for both materials.
- From the comparison of the simple mean and weighted mean presented in Appendix A with the RV for all GB, we observe that both means are always pretty close to the RV. We do not identify either any systematic effect between the Interferometric mean or RV and the mechanical comparison mean.
- Once Expanded Uncertainties are considered, the performance of most of the participants for the steel GB was good, even though three results had normalized errors greater than one: NPLI on the 75 mm, CMI on the 50 mm and CENAMEP on the 5 mm; and BSJ did not present results for the 1.000 5 mm GB.
- IBMETRO presented inconsistent results for the ceramic GB. As they obtained consistent results for the steel GB, we presumed their “raw” measurements of the Ceramic GB were probably good, but that they applied a wrong correction for comparing GB of two different materials. In effect, they informed us after circulation of DRAFT A, they made such a mistake and they are already amending the miscalculation in the procedures of their Quality Management System.
- Also for the ceramic GB TTBS obtained inconsistent results for the three longer ones; and TTBS and BSJ did not present results for two nominal values.
- As it has been discussed, some of the claimed calibration uncertainties seem too large for the calibration of lower grade GB. We suggest these laboratories review their uncertainties according to the grade of GB they aim to calibrate. In particular and with the present quoted uncertainties, BSJ would only be able to calibrate grade 1 and 2 GB and TTBS only grade 2.
- For the rest of the participants their results are judged satisfactory which proves their technical competence.

12. Acknowledgements

We would like to acknowledge:

- SIM WG.4 Length and SIM Technical Committee for having funded the purchase of the fourteen GB to carry this comparison.
- In an anonymous way, the technicians and colleagues from our different institutions that contributed directly or indirectly to the measurements of the artifacts in this comparison.

13. References

1. **ISO 3650:1998(E)**, *Geometrical Product Specification (GPS) – Length Standards – Gauge Blocks*, International Organization for Standardization, Geneva, Switzerland.
2. **C. Colín, M. Viliesid, et.al.** *SIM Regional Key Comparison SIM.L-K1:2007. Calibration of Gauge Blocks by Optical Interferometry*. June 2012.
3. **Thalmann R.**, *CCL-K1 Final report, Calibration of gauge blocks by interferometry*, Wabern, Switzerland, November 2000. At BIPM website, <http://www.bipm.fr>.

4. **Viliesid M.**, *Comparison CCL-K6 “Calibration of Coordinate Measuring Machine (CMM) Two-dimensional (2-D) Artifacts (Ball plates and Bore Plates)” Final Report*, 2008- 10-27 At BIPM website, <http://www.bipm.fr>.
5. **Cox, M.G.**, 2002, *The evaluation of key comparison data*, *Metrologia*, 2003, **39**, pp 589-595.
6. **Beissner, K.**, 2002, *On a measure of consistency in comparison measurements*, *Metrologia*, 2003, **39**, pp 59-63.
7. **Kacker, R., Datla, R., Parr, A.**, 2002, *Combined result and associated uncertainty from interlaboratory evaluations based on the ISO Guide*, *Metrologia*, 2003, **39**, pp 279-293.
8. **Decker J., Ulrich A., Lapointe A., Viliesid M., Pekelsky J. R.**, *Two-part Study towards Lowest Uncertainty Calibration of Ceramic Gauge Blocks: Interferometry And Mechanical Comparison Techniques*, in *Recent Developments in Traceable Dimensional Measurements*, Proceedings of SPIE – The International Society for Optical Engineering, **Vol. 4401** presented at Munich, Germany, June 2001.
9. **Decker J., Pekelsky J. R.**, *Uncertainty of Gauge Block Calibration by Mechanical Comparison: A Worked Example – Case 1: Gauges of Like Materials*, NRC Doc 39998, presented at National Conference of Standards Laboratories, Ottawa, Ontario, 1996.

Annex A Calculation of Alternate Statistical Parameters.

Steel gauge blocks / Nominal length, mm							
Statistical estimator	1.000 5 ¹	5	7	10	50 ²	75 ³	100 ⁴
Reference Value, RV	-9.3	26.4	-12.5	36.2	4.9	-105.6	-42.0
Standard uncertainty	4.1	4.2	4.2	3.9	6.3	8.5	8.6
Simple arithmetic mean	-5.9	20.9	-11.1	27.7	0.1	-104.6	-50.3
Standard uncertainty	7.7	7.6	7.8	8.7	11.7	17.1	17.6
Birge Ratio	0.77	0.83	0.83	0.57	0.40	0.50	0.54
Weighted mean	-10.2	16.9	-8.8	29.7	5.4	-104.9	-54.9
Standard uncertainty	5.5	5.4	5.5	5.7	8.6	10.5	13.8
Birge Ratio	0.87	0.97	0.94	0.76	0.44	0.62	0.56
Median	-11.0	10.5	-12.5	22.0	0.0	-109.5	-60.0
Observed chi-squared	10.7	17.3	14.6	10.1	2.6	5.0	4.6
Degrees of freedom	14	15	15	15	13	13	12
$Pr\{\chi^2(v) > \chi_{obs}^2\}$	0.71	0.30	0.94	0.82	1.00	0.976	0.97
Reduced chi-squared	0.76	1.16	0.97	0.67	0.20	0.38	0.39

¹ BSJ not considered in the statistical parameter calculations as they did not measure.

² CMI and TTBS were eliminated from the statistical parameter calculations.

³ NPLI and BSJ were eliminated from the statistical parameter calculations.

⁴ CMI, INEN and TTBS were eliminated from the statistical parameter calculations.

Ceramics gauge blocks / Nominal length, mm							
Statistical estimator	1.000 5 ⁵	5 ⁶	7 ⁶	10 ⁷	50 ⁸	75 ⁹	100 ⁸
Reference Value, RV	-3.9	10.6	51.7	-14.3	105.9	136.2	-23.1
Standard uncertainty	3.9	4.0	4.0	4.1	5.7	7.0	7.8
Simple arithmetic mean	-11.6	-2.0	33.9	-26.7	96.2	131.2	-13.0
Standard uncertainty	8.2	8.2	8.2	8.0	11.2	13.7	16.3
Birge Ratio	0.78	0.89	0.82	0.75	0.87	0.56	0.66
Weighted mean	-16.1	-0.5	39.8	-20.6	102.8	132.4	-17.5
Standard uncertainty	5.8	5.7	5.7	5.9	8.6	11.1	13.6
Birge Ratio	0.64	0.79	0.86	0.77	0.91	0.62	0.73
Median	-11.0	-10.0	34.0	-19.0	100.0	131.0	-5.0
Observed chi-squared	9.8	12.5	14.7	8.9	10.1	4.4	6.6
Degrees of freedom	13	14	14	13	12	11	12
$Pr\{\chi^2(\nu) > \chi_{obs}^2\}$	0.71	0.57	0.40	0.78	0.61	0.96	0.88
Reduced chi-squared	0.75	0.89	1.05	0.69	0.84	0.40	0.55

⁵ IBMETRO and BSJ were not considered for the statistical parameter calculations.

⁶ IBMETRO was not considered for the statistical parameter calculations.

⁷ IBMETRO and TTBS not considered for the statistical parameter calculations.

⁸ IBMETRO, TTBS and BSJ not considered for the statistical parameter calculations.

⁹ NPLI, IBMETRO, TTBS and BSJ were eliminated from the statistical parameter calculations.